

U.S. Army Medical Materiel Agency M3D Directorate Medical
Maintenance Operations Division, Tracy California
External Standing Operating Procedures

MCMR-MMO-MR

June 2011

1. Purpose

To provide guidance to units and organizations requesting services from USAMMA's Medical Maintenance Operations Division-California (MCMR-MMO-MR) at Defense Distribution Center, Tracy Location, Tracy CA 95304-9150.


2. Scope

These procedures are applicable to all units and activities requesting support.

3. Mission

The USAMMA Medical Maintenance Operations Division, Tracy, provides depot-level services and functions in support TO&E medical equipment specializing in x-ray equipment and Special Purpose Test Measurement Diagnostic Equipment (TMDE-SP).

3.1 Tracy serves as the regional manager for all medical TO&E except optical and audiometer equipment. The map below depicts Tracy's Region.



4. Hours of Operation

Normal duty hours are 0450 to 1520 (PT) daily Monday through Friday, excluding holidays. A telephone recorder is available on DSN 462-4557 or commercial (209) 839-4557 and will record messages after duty hours. When leaving messages, please speak clearly so your message will be understood. If you require a response, leave your name, telephone number, and the work order number if available. Recorded calls will be responded to on the following workday.

POSITION	COMMERCIAL	DSN
Chief	209 839-4556	462-4556
Shop Supervisor	209 839-4560	462-4560
Production Control	209 839-4557	462-4557
Fax	209 839-4563	462-4563
Website	http://www.usamma.army.mil/M3D.cfm	

5. Services Provided

5.1 Assistance Visits: On-site assistance visits will be conducted annually by Tracy Division for National Guard supported units within our region. This will be accomplished by division maintenance teams or arranged maintenance support with other maintenance activities in the state. The Chief/Shop Supervisor, Tracy Division, will coordinate scheduling of visits. All other assistance visits to include On-Site technical training and X-ray acceptance inspection requests will be coordinated through HQ, USAMMA. Please contact the Chief of the Medical Maintenance Division prior to submitting any request for assistance. Any unit desiring an on-site assistance visit, with the exception of National Guard units, will be required to submit a memorandum to:

MCMR-MMO-M
Mr. Jack Rosarius
Medical Maintenance Operations Division
693 Neiman Street
Fort Detrick, Maryland 21702-5001

5.2. Depot level services for all TO&E except optical and audiometer equipment.

5.3. Telephonic technical assistance.

5.4. Repair and Return services all TO&E except optical and audiometer equipment.

5.5. Tracy services all Army activities for calibrate and repair special purpose test, measurement, and diagnostic equipment (TMDE-SP) Listed in table below. If listed as Verify/OEM, Tracy will verify and if within calibration will complete. If item is out of calibration and needs adjustment, the item will be sent to OEM for repair and calibration. If listed as Calibrate/OEM Tracy will make adjustments to bring item within calibration. If unable to bring within tolerance Tracy will send to OEM for repair and calibration. To maintain capability when TMDE-SP is turned in for repair or calibration a like item may be borrowed, if available.

	Item/ Nomenclature	NSN	Model	Service
T.M.D.E. SP (Special Purpose)	Meter, X-Ray Calibration X-ray Calibration & Verification System	6525-01-502-0504 6625-01-312-0894	UNFORS 710-L Victoreen 07/457/472/473	Verify/OEM Calibrate
	Gas Flow Analyzer Calibrator, Gas Flow	6515-01-491-6615 6695-01-255-2855	VT-Plus/ VT-Plus HF RT-200	Calibrate/OEM Calibrate/OEM
	Anesthetic Gas Analyzer Anesthetic Digital Gas Analyzer	6630-01-487-6987 6630-01-530-7959	Riken 1802D Riken FI-21	Verify/OEM Verify/OEM
	Analyzer, NIBP	6515-01-449-1423	Cufflink	Calibrate/OEM
	IV Pump Analyzer	6515-01-449-2331 6515-01-479-2355 6515-01-536-2122	IPT-1 IDA-4/P3 IDA-4/P1	Calibrate Calibrate/OEM Calibrate/OEM
	Defibrillator Analyzer Tester, Defibrillator	6515-01-499-1420 6515-01-575-2825 6625-00-433-9063	Impulse 4000 Impulse 7000 DT2000A	Calibrate/OEM Calibrate/OEM Calibrate
	Densitometer, SU150/P	6525-01-161-1945	07-423	Calibrate
	Simulator, Medical Function	6625-01-298-3830 6625-01-286-2741	215M 217A	Calibrate Calibrate/OEM
	Calibrator Generator, ECG Multi-parameter Simulator	6515-01-049-9449 6625-01-548-3352	ECG 100 MPS-450	Calibrate Calibrate/OEM
	Foot Candle Meter Photometer	6695-01-303-0294 6695-01-574-8935	9-118 07-621	Calibrate Calibrate/OEM
	Thermometer, Dig.	6685-01-292-7873	51-II	Calibrate/OEM
	Wattmeter, Ultrasound Therapy	6625-01-504-2654 6625-01-141-7357 6625-01-487-6986 6625-01-540-8680 6695-01-575-6270	UW-4 UMR3-C UMR3-D UW-5 PUP 50	Calibrate/OEM Calibrate Calibrate Calibrate/OEM Calibrate/OEM
	Simulator, Pulse Oximeter	6515-01-504-8537 6515-01-518-4101 6515-01-499-1422	INDEX 2MFE INDEX 2XLFE Cardiosat	Calibrate/OEM Calibrate/OEM Calibrate
	Tachometer, Stroboscopic	6680-01-307-6190 6680-00-243-9977	1893A 1726	Calibrate Calibrate/OEM
	Test Set, Electro-surgical	6515-01-438-2409 6625-01-042-8213 6515-01-564-8554	454A RF302 QA-ES	Calibrate/OEM Calibrate Calibrate/OEM
	Tester, Current Leakage	6625-01-142-8233 6625-01-207-8270 6625-01-577-6744	232M 232D ESA 612	Calibrate Calibrate/OEM Calibrate/OEM
	Tester, Ventilator	6515-01-449-1421	Pneuvew 3600i	Calibrate/OEM

5.6. Direct Exchange of X-ray Tube Heads

An exchange for x-ray tubes may be requested by calling customer assistance at DSN 462-4560/4556 or commercial (209) 839-4560/4556. A questionnaire will be faxed to your activity to determine the appropriate information for the exchange. Tracy division rebuilds or provides a direct exchange for the following types of x-ray tube heads:

National Stock Number (NSN)	Tube Type
6525-00-C95-9590	Picker PX1300 Series
6525-00-C95-9600	Picker PX1400 Series
6525-00-C95-9560	Eureka Emerald 125 Series
6525-00-C95-9550	Eureka Sapphire 150 th Series
6525-00-C95-9570	Eureka Diamond 150 th Series
6525-00-C95-9785	Philips ROT 350-10
6525-00-C95-9630	G.E. Maxiray 75
6525-00-C95-9640	G.E. Maxiray 100
6525-00-C95-9640	G.E. Maxiray 100 Flouro
6525-01-328-3430	Dynarad/Port-A-Ray 1200

6. Requesting Services

6.1. Prior to sending any nonstandard medical equipment, call DSN 462-4557 to ensure that the items can be supported at this division.

6.2. When shipping equipment for repair or service, please use the following address:

Medical Maintenance Operations Division
 USAMMA, Building T-255
 25600 South Chrisman Road
 Tracy, CA 95304-9150 DODAAC: W62SEV

6.3. All units, organizations, facilities or agencies other than Active Army (P84 and medical P1 funded) are required to reimburse USAMMA for all services. Army National Guard and U.S. Army Reserve units are not required to submit fund citations as their respective headquarters provide funds on an annual basis to cover their medical equipment. Funding documentation from other reimbursable customers must include the following:

* Document number to include owning DODAAC or UIC, and address
* Funding citation
* Authorized funding (amount authorized for service)
* Point of contact and telephone number
* Nomenclature of item
* National stock number, management control number, or non-standard number
* Quantity of items to include serial numbers
* Any accessories, maintenance manuals, or other materiel which may be required to perform services on the equipment
* Identification of all accessories

6.4. Questions concerning funding or fund citations may be answered by calling DSN 462-4557 or commercial 209 839-4557.

6.5. All customers may request maintenance services on their medical equipment by submitting either a DA Form 2407 (or the automated equivalent) or DD Form 1348-1, Shipping Document. TMDE-SP equipment must include DA7372, POC and a return address.

6.6. All equipment that has reusable containers will be shipped in those containers. If equipment does not have reusable container equipment will be packed so that no further damage can occur.

6.7. Place a copy of the document being used as the maintenance request inside the shipping container with the equipment. The transportation personnel or the commercial carrier often removes documents placed on the outside of the container.

6.8. Accessories and maintenance manuals must be sent with the equipment to prevent delays in the repair or service. All accessories or materials sent with the equipment shall be indicated in the remarks section of the DA Form 2407 and DD Form 1348-1 or by other documentation.

6.9. The Maintenance Expenditure Limit (MEL) shall be included in the remarks section of either the DA Form 2407 or DD Form 1348-1. Failure to include the MEL will result in delay of repairs.

6.10. When Active Army units submit equipment to the MMOD-Tracy that belongs to another unit, the owning unit, address, UIC, and DODAAC will be given in the remarks section of either the DA Form 2407 or DD Form 1348-1. Unless otherwise specified, after repairs are completed the equipment will be returned to the owning unit.

6.11. Equipment items not listed in 5. Services or on the USAMMA Maintenance website should not be sent without prior coordination.

6.12. The USAMMA Medical Maintenance Operations Division-California (MMOD-Tracy) is not responsible for billing reimbursable customers. For questions concerning billing call USAMMA's Resources Management Division at DSN 343-2111 or commercial 301-619-2111.

6.13. Any questions regarding MMOD-Tracy's services, work order status, complaints, technical assistance or general information may be answered by calling DSN 462-4557/4556/4560 or commercial 209 839-4557/4556/4560. Please have the serial number of equipment item and work order number available when you call.

7. Direct Exchange of X-ray Tube Heads

An exchange for x-ray tubes may be requested by calling customer assistance at DSN 462-4560/4556 or commercial (209) 839-4560/4556. A questionnaire will be faxed to your activity to determine the appropriate information for the exchange.

8. Medical Standby Equipment Program

The USAMMA MMOD-Tracy provides a Medical Standby Equipment Program (MEDSTEP) for selected x-ray equipment. A list of the MEDSTEP assets available at the MMOD-CA is published periodically in the SB 8-75 Series Bulletins. MEDSTEP assets may only be utilized to provide serviceable temporary replacement for equipment being serviced at the MMOD-Tracy. The USAMMA Maintenance Operations Division must approve exceptions. Exceptions may be requested telephonically by calling DSN 343-4365 or commercial 301-619-4365. Once the owner's original equipment is received back, the MEDSTEP item, to include all accessories, must be returned to the MMOD-CA. Reimbursable customers that use MEDSTEP must provide funds as necessary to restore the MEDSTEP item back to serviceable condition.

9. Cannibalization Point

The USAMMA MMOD-Tracy maintains unserviceable assets of selected medical equipment for cannibalization. Authorized customers may request parts from Cannibalization for mission critical medical equipment when parts are not available from any other source.

10. Training

Repair and calibration training of various x-ray equipment items is available. Training is tailored to meet your requirements. Training can be provided either in your facility or in the Division facilities. On-site training costs are based upon travel, per diem, the labor rate and training hours. Contact us about your needs at DSN 462-4556 or commercial (209) 839-4556.

Chief, Medical Maintenance
Operations Division
USAMMA