

Table, Operating, Hospital, Model 2080, NSN 6530-01-353-9883

The operating table, manufactured by Steris Corporation, LIN T00029, is supplied with a number of accessory components. The list of accessories supplied with the operating table (listed in table 1-3) is taken from the Medical Procurement Item Description (MPID).

For ease of inventory and operational readiness, you should make a copy of this table, laminate it in plastic, and include it with the operator and service documentation.

Table 1-3. Required Accessories for the Operating Table

Part Number	Item Description
18688-091	Holder (pair)
77033-091	Ether Screen Assembly
77035-091	Lateral Braces (pair)
77036-091	Wrist Holder Assembly (pair)
77040-091	Lithotomy Legholders (pair)
83594-001	Pads (listed below)
P093074-001	Pad, Foot Section
P093075-001	Pad, Back Section
P093076-001	Pad, Head Section
129357-096	X-ray Top Sections (4 each) Section 1 - Head Section 2 - Back Section 3 - Seat Section 4 - Leg
P056397-500 (Old Number) P150830-168 (New Number)	3" Arm Board Pad
56130-001	I.V. Arm Board Without Pad
B146645-277	62" Image Intensifier Board
No Part Number	Clark Sockets (4 each)

For more information on parts, write Steris Corporation at 2424 W. 23rd Street, Erie, PA 16506, or call 800-333-8828.